

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/22/2022

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:32
INTERNAL STANDARD	VERIFIED	16:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:32
BLANK TEST	0.00	16:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:33
BLANK TEST	0.00	16:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:35
BLANK TEST	0.00	16:35

Sgt. Charles B. Geisler #244

