

SOLUTION CHANGE

WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 105208

Date: 05/02/2022

OPERATOR NAME:  
 JOSHUA M FISH  
 Permit #: 37493  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:49
INTERNAL STANDARD	VERIFIED	14:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:49
BLANK TEST	0.00	14:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:50
BLANK TEST	0.00	14:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:52
BLANK TEST	0.00	14:52

*PO #427*

