

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/16/2022

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:02
INTERNAL STANDARD	VERIFIED	09:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:03
BLANK TEST	0.00	09:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:04
BLANK TEST	0.00	09:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:05
BLANK TEST	0.00	09:06



Handwritten signature and number 714