

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/23/2022

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:20
INTERNAL STANDARD	VERIFIED	01:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:21
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:22
BLANK TEST	0.00	01:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:23
BLANK TEST	0.00	01:24



[Handwritten signature]
#820