

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/15/2022

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN COUNTY SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:30
INTERNAL STANDARD	VERIFIED	00:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:31
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:32
BLANK TEST	0.00	00:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:33
BLANK TEST	0.00	00:34

J. T. Coon

