

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/23/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:21
INTERNAL STANDARD	VERIFIED	09:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:21
BLANK TEST	0.00	09:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:23
BLANK TEST	0.00	09:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:24
BLANK TEST	0.00	09:25

Steven T Bunio #789

