

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 07/01/2022

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:26
INTERNAL STANDARD	VERIFIED	10:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:27
BLANK TEST	0.00	10:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:28
BLANK TEST	0.00	10:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:29
BLANK TEST	0.00	10:30

*PO Steven Bunio #387*

