

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 07/29/2022

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:10
INTERNAL STANDARD	VERIFIED	10:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:10
BLANK TEST	0.00	10:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:11
BLANK TEST	0.00	10:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:13
BLANK TEST	0.00	10:13

*PO Fish #427*

