

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/22/2022

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:37
INTERNAL STANDARD	VERIFIED	10:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:38
BLANK TEST	0.00	10:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:39
BLANK TEST	0.00	10:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:40
BLANK TEST	0.00	10:41

P.O. *[Signature]* #301

8/22/2022

