

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/29/2022

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:48
INTERNAL STANDARD	VERIFIED	03:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:48
BLANK TEST	0.00	03:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:50
BLANK TEST	0.00	03:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:51
BLANK TEST	0.00	03:52

