

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/01/2022




OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:28
INTERNAL STANDARD	VERIFIED	15:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:28
BLANK TEST	0.00	15:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:30
BLANK TEST	0.00	15:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:31
BLANK TEST	0.00	15:32

Handwritten signature # 301

9/02/22



SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/01/2022

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:10
INTERNAL STANDARD	VERIFIED	08:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:10
BLANK TEST	0.00	08:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:11
BLANK TEST	0.00	08:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:12
BLANK TEST	0.00	08:13

Kenneth W Smith #301

9/01/2022

