

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/09/2022

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:31
INTERNAL STANDARD	VERIFIED	17:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:32
BLANK TEST	0.00	17:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:33
BLANK TEST	0.00	17:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:34
BLANK TEST	0.00	17:35

[Handwritten signature]

9/9/2022

