

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/03/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:06
INTERNAL STANDARD	VERIFIED	17:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:07
BLANK TEST	0.00	17:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:08
BLANK TEST	0.00	17:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:09
BLANK TEST	0.00	17:10

PT Steven T Bunio #189

