

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/09/2022

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:48
INTERNAL STANDARD	VERIFIED	08:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:48
BLANK TEST	0.00	08:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:49
BLANK TEST	0.00	08:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:50
BLANK TEST	0.00	08:51

POA #4137

