

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/24/2022

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:27
INTERNAL STANDARD	VERIFIED	02:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:28
BLANK TEST	0.00	02:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:29
BLANK TEST	0.00	02:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:30
BLANK TEST	0.00	02:31

*H 820*

