

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/14/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:32
INTERNAL STANDARD	VERIFIED	11:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:32
BLANK TEST	0.00	11:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:33
BLANK TEST	0.00	11:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:34
BLANK TEST	0.00	11:35

PO *Steven Bunio* #789

