

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/21/2022

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:46
INTERNAL STANDARD	VERIFIED	03:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:46
BLANK TEST	0.00	03:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:47
BLANK TEST	0.00	03:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:49
BLANK TEST	0.00	03:49

*Michael A Campbell*  
#157

