

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/28/2022

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:11
INTERNAL STANDARD	VERIFIED	10:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:11
BLANK TEST	0.00	10:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:13
BLANK TEST	0.00	10:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:14
BLANK TEST	0.00	10:15

20 g/210L 11/27

