

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 01/10/2022

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:01
INTERNAL STANDARD	VERIFIED	12:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:02
BLANK TEST	0.00	12:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:03
BLANK TEST	0.00	12:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:04
BLANK TEST	0.00	12:05



*PO J. Fish #427*