

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/24/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:21
INTERNAL STANDARD	VERIFIED	08:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:22
BLANK TEST	0.00	08:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:24
BLANK TEST	0.00	08:25

PT Stev Bunio #257

