

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 02/08/2022

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:03
INTERNAL STANDARD	VERIFIED	10:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:03
BLANK TEST	0.00	10:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:04
BLANK TEST	0.00	10:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:06
BLANK TEST	0.00	10:06

*PT Steu #789*

