

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 02/16/2022

OPERATOR NAME:  
SEAN T SMITH  
Permit #: 46227  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:27
INTERNAL STANDARD	VERIFIED	10:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:27
BLANK TEST	0.00	10:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:29
BLANK TEST	0.00	10:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:30
BLANK TEST	0.00	10:31

