

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 02/23/2022

OPERATOR NAME:
JAMES RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:28
INTERNAL STANDARD	VERIFIED	19:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:28
BLANK TEST	0.00	19:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:30
BLANK TEST	0.00	19:31

[Handwritten signature]
#1820

