

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 03/03/2022

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:13
INTERNAL STANDARD	VERIFIED	00:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:14
BLANK TEST	0.00	00:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:15
BLANK TEST	0.00	00:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:16
BLANK TEST	0.00	00:17



*Michael A. Campbell #389*  
*Michael A. Campbell*