

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 03/08/2022

OPERATOR NAME:  
KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:24
INTERNAL STANDARD	VERIFIED	14:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:25
BLANK TEST	0.00	14:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:26
BLANK TEST	0.00	14:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:27
BLANK TEST	0.00	14:28

*20. Handwritten signature #301*

*3/8/2022*

