

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 03/15/2022

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:19
INTERNAL STANDARD	VERIFIED	13:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:20
BLANK TEST	0.00	13:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:21
BLANK TEST	0.00	13:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:22
BLANK TEST	0.00	13:23

