

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/22/2022

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:16
INTERNAL STANDARD	VERIFIED	10:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:17
BLANK TEST	0.00	10:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:18
BLANK TEST	0.00	10:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:19
BLANK TEST	0.00	10:20

PO Afil #407

