

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/31/2022

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:55
INTERNAL STANDARD	VERIFIED	14:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:56
BLANK TEST	0.00	14:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:58
BLANK TEST	0.00	14:59



P.C. [Signature] #714