

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 04/22/2022

OPERATOR NAME:
 JOSHUA M FISH
 Permit #: 37493
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:38
INTERNAL STANDARD	VERIFIED	15:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:39
BLANK TEST	0.00	15:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:40
BLANK TEST	0.00	15:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:41
BLANK TEST	0.00	15:42

PO #427

