

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 05/07/2022

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:21
INTERNAL STANDARD	VERIFIED	11:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:22
BLANK TEST	0.00	11:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:23
BLANK TEST	0.00	11:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:24
BLANK TEST	0.00	11:25

*PO #427*

