

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/16/2022

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:19
INTERNAL STANDARD	VERIFIED	08:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:20
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:21
BLANK TEST	0.00	08:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:22
BLANK TEST	0.00	08:23



P.O. [Signature] 714