

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 06/23/2022

OPERATOR NAME:
 STEVEN T BUNIO
 Permit #: 44983
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:34
INTERNAL STANDARD	VERIFIED	08:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:35
BLANK TEST	0.00	08:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:36
BLANK TEST	0.00	08:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:37
BLANK TEST	0.00	08:38

PO Steven Bunio # 789

