

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/01/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:55
INTERNAL STANDARD	VERIFIED	08:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:56
BLANK TEST	0.00	08:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:57
BLANK TEST	0.00	08:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:58
BLANK TEST	0.00	08:59

PO Steven T Bunio

