

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/07/2022

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:30
INTERNAL STANDARD	VERIFIED	09:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:30
BLANK TEST	0.00	09:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:31
BLANK TEST	0.00	09:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:33
BLANK TEST	0.00	09:33

PO Kenneth W. Smith #301

7/07/2022

