

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/16/2022

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:26
INTERNAL STANDARD	VERIFIED	07:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:26
BLANK TEST	0.00	07:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:28
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:29
BLANK TEST	0.00	07:30

PO [Signature] #427

