

SUPERVISOR TEST


WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/23/2022

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:42
INTERNAL STANDARD	VERIFIED	19:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:42
BLANK TEST	0.00	19:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:44
BLANK TEST	0.00	19:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:45
BLANK TEST	0.00	19:46



PO *[Signature]* #757