

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 08/22/2022

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:51
INTERNAL STANDARD	VERIFIED	08:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:51
BLANK TEST	0.00	08:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:53
BLANK TEST	0.00	08:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:54
BLANK TEST	0.00	08:55

*PO Steve Bunio 789*  
*8.22.22*

