

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 09/23/2022

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:36
INTERNAL STANDARD	VERIFIED	12:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:37
BLANK TEST	0.00	12:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:38
BLANK TEST	0.00	12:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:39
BLANK TEST	0.00	12:40

P.O. [Signature] #301

9/23/2022

