

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/14/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:47
INTERNAL STANDARD	VERIFIED	09:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:48
BLANK TEST	0.00	09:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:49
BLANK TEST	0.00	09:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:50
BLANK TEST	0.00	09:51

PO Steven Bunio #789

