

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/12/2022

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:10
INTERNAL STANDARD	VERIFIED	08:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:11
BLANK TEST	0.00	08:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:12
BLANK TEST	0.00	08:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:14

20412#427

