

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/10/2022

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:05
INTERNAL STANDARD	VERIFIED	22:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:07
BLANK TEST	0.00	22:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:09

10/1/2011

