

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 02/16/2022

OPERATOR NAME:  
SEAN T SMITH  
Permit #: 46227  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:34
INTERNAL STANDARD	VERIFIED	11:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:34
BLANK TEST	0.00	11:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:36
BLANK TEST	0.00	11:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:37
BLANK TEST	0.00	11:38

