

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/02/2022

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:30
INTERNAL STANDARD	VERIFIED	21:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:30
BLANK TEST	0.00	21:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:31
BLANK TEST	0.00	21:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:33
BLANK TEST	0.00	21:34



PO M.A. CAMPBELL #757
PO M.A. CAMPBELL