

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/08/2022

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:22
INTERNAL STANDARD	VERIFIED	07:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:24
BLANK TEST	0.00	07:25

P.O. Smith #301

3/8/2022

