

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/15/2022

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:20
INTERNAL STANDARD	VERIFIED	14:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:21
BLANK TEST	0.00	14:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:22
BLANK TEST	0.00	14:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:23
BLANK TEST	0.00	14:24

Donnie J Long
844

