

SUPERVISOR TEST




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/23/2022

OPERATOR NAME:  
J T COON  
Permit #: 36043  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:03
INTERNAL STANDARD	VERIFIED	01:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:04
BLANK TEST	0.00	01:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:05
BLANK TEST	0.00	01:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:06
BLANK TEST	0.00	01:07



*P. J. Coon* 341