

SOLUTION CHANGE




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/31/2022

OPERATOR NAME:  
JOHN D BATEHOLTS  
Permit #: 43007  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:26
INTERNAL STANDARD	VERIFIED	16:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:27
BLANK TEST	0.00	16:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:28
BLANK TEST	0.00	16:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:29
BLANK TEST	0.00	16:30



*John D Bateholts* 6714