

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/08/2022

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:26
INTERNAL STANDARD	VERIFIED	15:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:26
BLANK TEST	0.00	15:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:27
BLANK TEST	0.00	15:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:28
BLANK TEST	0.00	15:29

