

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/22/2022

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:30
INTERNAL STANDARD	VERIFIED	12:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:31
BLANK TEST	0.00	12:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:32
BLANK TEST	0.00	12:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:33
BLANK TEST	0.00	12:34

DOA #427

