

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/01/2022

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:04
INTERNAL STANDARD	VERIFIED	09:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:05
BLANK TEST	0.00	09:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:06
BLANK TEST	0.00	09:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:07
BLANK TEST	0.00	09:08

*Donnie J Long*  
844

